

GREYWELL PARISH COUNCIL

Risk Assessment for Defibrillator

	Task Element	Hazard and harmful effect	Initial risk level			Control measures to reduce likelihood	Controlled risk level to Individual		
			Severity	Likelihood	Risk level		Severity	Likelihood	Risk level
1	Defibrillator is used by untrained personnel.	Defibrillator is not used appropriately.	3	2	5 H	The Defibtech DDU 2000 series defibrillator has a video screen and shouts commands to the rescuer. It is designed to be used by both trained and untrained operators.	1	2	3 M
2	The defibrillator is not functioning.	The defibrillator is malfunctioning or the battery has expired.	3	1	3 M	The defibrillator (and battery status) is given a cursory check at least once a week and a full check every 4 weeks. In the case of continued malfunction, the village will be informed by the village email system and the Community Heartbeat Trust will inform the local ambulance service. Any attending ambulance will have a defibrillator.	1	1	2 L

Adopted March 2024

Review Date March 2025